


DATE	SYM	REVISION	RECORD	AUTH.	DR.	CK.
8/3/99	-	ISSUED			NT	DG



FIRST END— SAFETY HOOK:

GOLD PLATED BERYLLIUM COPPER CONTACT

LEXAN INSULATION

WIRE: 20 AWG 41/36 TC PVC JACKET.087"O.D.

SECOND END— BANANA PLUG:

NICKEL PLATED BRASS BODY

NICKEL PLATED BERYLLIUM COPPER SPRING

POLYPROPYLENE INSULATION

PART NUMBER:

WIRE LENGTH:

795-5025-XX-25

12"

795-5026-XX-25

24"

795-5027-XX-25

36"

795-5028-XX-25

48"

795-5029-XX-25

60"

TOLERANCES (EXCEPT AS NOTED)	CONCORD		ELECTRONICS CORPORATION
DECIMAL + - .010	P/N	SCALE	30 GREAT JONES ST. NEW YORK N. Y. 10012
FRACTIONAL + -	TITLE SAFETY HOOK TO IN-LINE BANANA PLUG		
ANGULAR + -	DATE 8/3/99	DRAWING NUMBER 795-5025-XX-25_5029-XX-25	